

Serix White – Nitrile Examination Gloves

Product Specs



Description

- Serix White disposable 100% nitrile examination gloves are manufactured at the highest quality standards & offer increased protection and excellent chemical resistance
- Stronger against punctures and tears than latex gloves
- Easy to put on and take off thanks to their coefficient of friction
- Due to their close-fitting design and finger textured, they can be used with high dexterity, including wet environments
- Serix White Nitrile Gloves are intended to be worn on the hands of healthcare personnel during medical examination procedure to protect cross contamination between healthcare personnel and patient.
- Tested and certified as a medical device according to EN 455

Areas of use

- Hospitals, Laboratories, Clinics
- Health Care
- Dentistry and veterinary applications
- Food production and processing
- Cosmetics Industry

Advantages

- Ambidextrous and feature a tight-fitting design and textured fingers that provides better movement and handling
- Suitable for individuals with latex sensitivity/allergies
- Ideal for use in food processing and food service
- Super soft and comfortable.
- Multilanguage text on box (EN, RO, DE, IT, ES, CZ, PL, GR, BG)

Specifications

Manufacturer	Blue Sail Medical Co., Ltd				
EC-REP	Lotus NL B.V				
Brand	Serix White				
Color	White				
Material	100% Nitrile, Powder Free				
Length	min. 240 mm				
Thickness	0,07 mm. (fingers) 0,06 mm. (palm)				
Size	XS	S	M	L	XL
Palm width (mm) (± 10)	70	80	95	110	120
Glove weight (g) (±0,3)	2,8	3,0	3,5	4,0	4,5
Resistance	6 N				
AQL Value 1,5	Yes				
Textured	Fingers				
Packaging	1 ctn = 100 gloves x 10 boxes				
Box size	210 x 115 x 60 mm				
Sterilization	Non-Sterile, Single Use Only				
Medical Device	Yes, Class I				
Ambidextrous	Yes				

Certifications and Standards

EN 455 - 1, 2, 3, 4
EN ISO 13485:2016
ASTM D6319-19, 6124-06 (2017)
EN ISO 374-5: 2016, EN374-4:2013
EN ISO 374-1: 2016/Type B (KPT)
EN 420: 2003 + A1:2009 (Palm Protection)
CE2777 comply to PPE Regulation 2016/425 Cat III

